



CANADA STATEMENT OF CONFORMITY

This Statement of Conformity is issued under the sole responsibility of the manufacturer.

MANUFACTURER	
Name of Company	Address
Delta Development Team INC	1635 S Research Loop #303 Tucson, Arizona 85710 USA
PRODUCT IDENTIFICATION	
Device Name	Device Model
Autonomous Portable Refrigeration Unit (APRU)	6L-2
INTENDED PURPOSE	
Blood Refrigerator	NOTE: In Canada, blood bank refrigerators do not meet the definition of “device” per the Food and Drugs Act R.S.C., 1985, c. F-27 (i.e. it is not represented for use in diagnosing / treating / mitigating / preventing a disease, restoring / modifying / correcting a body structure or functioning of a body part). Therefore, are not regulated under the Medical Devices Regulations. CAN/CSA-Z902:20 applies
Standards	
<p>CERTIFIED TO:</p> <ul style="list-style-type: none"> CSA C22.2 61010-1 2010 Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use CSA C22.2 61010-2-11 2019 Particular requirements for refrigerating equipment CSA C22.2 61010-2-10 2019 Particular requirements for laboratory equipment for the heating of materials <p>COMPLIES WITH;</p> <ul style="list-style-type: none"> CAN/CSA-Z902:20 Blood Standard. Sections 6.3, Table 2, 9.4, 23.1.3(c) per Blood Regulations SOR/2013-178 IEC 61326-1: 2013 Electrical equipment for measurement, control and laboratory-EMC; General IC RSS-247 Issue 2. Digital Transmission Systems IC RSS-102 Radio Frequency (RF) Exposure Compliance. Bluetooth SOR/2016-137 Halocarbon Regulations. Schedule 1.1 Subsections 64.4(1) and (2). Commercial Mobile Refrigeration System. HFC Limit less than GWP 2200 	

COMPANY REPRESENTATIVE: William Barg

TITLE: Chief Engineer

PLACE: 1635 S Research Loop #303 Tucson AZ USA

SIGNATURE:



DATE: 25 August 2022

Page 1 of 1